
Midmark M11 - v1.0.0
Total Cycles: 2391

Name: RITTER 1
Sterilizer ID: V2178198
10/03/2022 03:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.3 F 0.0
2:00 192.5 F 3.7
4:00 203.9 F 8.5
6:00 224.2 F 9.8
8:00 234.2 F 9.5
10:00 236.7 F 8.7
12:00 245.0 F 13.5
14:00 244.5 F 12.5
16:00 257.5 F 19.4
18:00 269.0 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.2 F 28.7
2:00 272.1 F 28.9
3:00 272.0 F 28.5
4:00 272.2 F 29.1

Min 271.5 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.9 F 0.0
5:00 215.8 F 0.0
10:00 223.3 F 0.0
15:00 224.0 F 0.0
20:00 222.9 F 0.0
25:00 222.3 F 0.0
30:00 220.8 F 0.0
Drying Complete

FILLING: 2:41
HEATING: 19:04
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:56
CYCLE COMPLETE

10/03/2022 04:54 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____