
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001263
11:23 21/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:44
PRESSURIZING 2:51
STERILIZING 3:57
136.0 C 220kPa 3:57
136.4 C 225kPa 4:27
136.2 C 222kPa 4:57
136.3 C 224kPa 5:27
136.4 C 224kPa 5:57
136.4 C 224kPa 6:27
136.4 C 224kPa 6:57
136.4 C 224kPa 7:27
136.4 C 224kPa 7:27
VENTING 7:27
AIR DRYING 8:17
CYCLE COMPLETE 18:51

Digital Signature #
CF6CEB11A16509B3
