

-----  
Midmark M11 - v1.0.5  
Total Cycles: 852

Name: RITTER2  
Sterilizer ID: V2313228  
03/01/2022 01:01 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 158.1 F 0.0  
2:00 157.5 F 2.9  
4:00 179.1 F 7.0  
6:00 202.4 F 12.1  
8:00 220.4 F 6.4  
10:00 232.6 F 11.9  
12:00 234.8 F 9.4  
14:00 237.9 F 9.5  
16:00 248.8 F 15.2  
18:00 261.4 F 22.4  
20:00 272.8 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.0  
1:00 272.7 F 29.4  
2:00 271.3 F 29.1  
3:00 272.6 F 29.4  
4:00 271.2 F 28.9

Min 271.1 F 28.8  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.0 F 0.0  
5:00 229.4 F 0.0  
10:00 216.9 F 0.0  
15:00 196.7 F 0.0  
20:00 194.5 F 0.0  
25:00 191.2 F 0.0  
30:00 190.0 F 0.0  
Drying Complete

FILLING: 1:23  
HEATING: 20:19  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:58:52  
CYCLE COMPLETE

03/01/2022 02:38 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02