
Midmark M11 - v1.0.5
Total Cycles: 2954

Name: RITTER2
Sterilizer ID: V2313228
10/02/2023 08:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.3 F 0.0
2:00 71.7 F 0.0
4:00 76.2 F 0.0
6:00 95.3 F 0.0
8:00 139.6 F 0.0
10:00 174.4 F 0.6
12:00 195.3 F 2.7
14:00 209.6 F 5.9
16:00 220.1 F 9.1
18:00 228.0 F 12.9
20:00 230.5 F 8.3
22:00 241.3 F 13.1
24:00 239.9 F 10.8
26:00 237.1 F 9.2
28:00 250.7 F 15.7
30:00 261.5 F 21.9
32:00 271.4 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.4 F 28.9
2:00 271.9 F 29.0
3:00 272.0 F 28.6
4:00 272.4 F 29.3

Min 271.5 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.7 F 0.0
5:00 208.2 F 0.0
10:00 201.4 F 0.0
15:00 188.4 F 0.0
20:00 186.7 F 0.0
25:00 207.2 F 0.0
30:00 208.6 F 0.0

Drying Complete

FILLING: 0:49
HEATING: 32:38
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:10:48
CYCLE COMPLETE

10/02/2023 09:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____