
Midmark M11 - v1.0.0
Total Cycles: 3625

Name: RITTER 1
Sterilizer ID: V2178198
08/09/2023 11:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 220.1 F 0.0
2:00 199.8 F 3.0
4:00 196.3 F 7.5
6:00 223.6 F 12.7
8:00 227.1 F 7.8
10:00 239.2 F 12.9
12:00 238.2 F 10.0
14:00 236.4 F 8.5
16:00 248.1 F 14.3
18:00 259.4 F 20.5
20:00 269.6 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.3 F 29.1
2:00 271.5 F 28.4
3:00 271.8 F 28.9
4:00 272.1 F 28.9

Min 271.1 F 28.2
Max 272.4 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.2 F 0.0
5:00 225.3 F 0.0
10:00 240.0 F 0.0
15:00 228.8 F 0.0
20:00 223.4 F 0.0
25:00 221.4 F 0.0
30:00 218.8 F 0.0
Drying Complete

FILLING: 1:41
HEATING: 21:03
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:00:02
CYCLE COMPLETE

08/09/2023 12:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____