
Midmark M11 - v1.0.0
Total Cycles: 1523

Name: RITTER 1
Sterilizer ID: V2178198
01/13/2022 11:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.2 F 0.0
2:00 168.1 F 3.7
4:00 192.1 F 7.8
6:00 215.1 F 12.6
8:00 223.4 F 6.4
10:00 233.9 F 11.4
12:00 232.9 F 8.0
14:00 243.8 F 12.0
16:00 244.3 F 12.4
18:00 256.6 F 18.9
20:00 267.9 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.8 F 29.1
2:00 271.5 F 28.3
3:00 271.9 F 29.0
4:00 272.4 F 29.0

Min 271.1 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.7 F 0.0
5:00 232.0 F 0.0
10:00 221.8 F 0.0
15:00 228.4 F 0.0
20:00 228.8 F 0.0
25:00 231.1 F 0.0
30:00 231.9 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 21:20
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:01
CYCLE COMPLETE

01/13/2022 12:57 PM
ACCEPTED: X
REJECTED:
Operator: U02