
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000836
15:47 04/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:50
PRESSURIZING 3:53
STERILIZING 5:08
136.0 C 226kPa 5:08
136.4 C 232kPa 5:38
136.4 C 232kPa 6:08
136.2 C 230kPa 6:38
136.3 C 231kPa 7:08
136.4 C 231kPa 7:38
136.4 C 232kPa 8:08
136.4 C 232kPa 8:38
136.4 C 232kPa 8:38
VENTING 8:38
AIR DRYING 9:35
CYCLE COMPLETE 24:35

Digital Signature #
C189E1C8143D9E32
