
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 003329
15:07 29/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:53
PRESSURIZING 2:58
STERILIZING 4:02
136.4 C 220kPa 4:02
136.2 C 223kPa 4:32
135.9 C 220kPa 5:02
136.4 C 224kPa 5:32
136.4 C 225kPa 6:02
136.1 C 221kPa 6:32
136.4 C 223kPa 7:02
136.4 C 219kPa 7:32
136.4 C 219kPa 7:32
VENTING 7:32
AIR DRYING 8:28
CYCLE COMPLETE 12:40

Digital Signature #
CC187C8847A8F8F3
