
Midmark M11 - v1.0.0
Total Cycles: 2171

Name: RITTER 1
Sterilizer ID: V2178198
07/25/2022 02:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.0 F 0.0
2:00 168.2 F 1.8
4:00 187.1 F 5.8
6:00 206.6 F 10.2
8:00 221.3 F 6.1
10:00 229.0 F 9.2
12:00 234.5 F 10.6
14:00 239.6 F 11.1
16:00 235.9 F 8.9
18:00 250.1 F 15.4
20:00 261.3 F 21.9
22:00 271.8 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.5
1:00 271.4 F 28.7
2:00 272.6 F 29.4
3:00 272.4 F 29.0
4:00 271.8 F 28.6

Min 271.2 F 28.3
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.1 F 0.0
5:00 231.1 F 0.0
10:00 232.0 F 0.0
15:00 230.8 F 0.0
20:00 229.2 F 0.0
25:00 228.3 F 0.0
30:00 226.5 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 22:30
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:38
CYCLE COMPLETE

07/25/2022 03:56 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____