
Midmark M11 - v1.0.5
Total Cycles: 4458

Name: RITTER2
Sterilizer ID: V2313228
03/26/2025 10:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.5 F 0.0
2:00 151.5 F 1.5
4:00 155.1 F 4.6
6:00 174.2 F 8.7
8:00 191.7 F 13.4
10:00 219.0 F 7.2
12:00 231.3 F 11.8
14:00 231.3 F 8.5
16:00 241.3 F 11.8
18:00 243.0 F 12.8
20:00 254.8 F 18.9
22:00 265.5 F 25.4
24:00 271.4 F 30.3

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 30.4
1:00 272.2 F 30.8
2:00 272.3 F 30.5
3:00 272.0 F 30.3
4:00 272.0 F 30.3

Min 271.7 F 30.3
Max 272.5 F 30.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.5 F 0.0
5:00 197.1 F 0.0
10:00 215.4 F 0.0
15:00 217.9 F 0.0
20:00 217.1 F 0.0
25:00 180.3 F 0.0
30:00 205.7 F 0.0
Drying Complete

FILLING: 1:27
HEATING: 24:26
STERILIZING: 4:00
VENTING: 3:54
DRYING: 30:00
TOTAL CYCLE: 01:03:48
CYCLE COMPLETE

03/26/2025 11:44 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____