

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3811

Name: RITTER2  
Sterilizer ID: V2313228  
07/24/2024 11:13 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 116.9 F 0.0  
2:00 116.2 F 0.7  
4:00 121.7 F 2.7  
6:00 141.2 F 5.5  
8:00 165.4 F 9.3  
10:00 184.6 F 13.5  
12:00 210.1 F 6.3  
14:00 225.0 F 10.6  
16:00 229.0 F 8.0  
18:00 238.7 F 12.1  
20:00 238.0 F 10.3  
22:00 250.0 F 16.2  
24:00 260.5 F 22.3  
26:00 270.7 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 30.0  
1:00 271.8 F 30.1  
2:00 271.8 F 30.1  
3:00 271.6 F 30.1  
4:00 271.8 F 29.9

Min 271.6 F 29.9  
Max 273.0 F 30.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.7 F 0.0  
5:00 217.2 F 0.0  
10:00 215.9 F 0.0  
15:00 225.0 F 0.0  
20:00 219.6 F 0.0  
25:00 220.8 F 0.0  
30:00 217.8 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 27:27  
STERILIZING: 4:00  
VENTING: 3:43  
DRYING: 30:00  
TOTAL CYCLE: 01:06:04  
CYCLE COMPLETE

07/24/2024 12:19 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_