

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 006817  
9:21 12/01/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:49  
PRESSURIZING 2:51  
STERILIZING 3:55  
136.4 C 220kPa 3:55  
136.1 C 220kPa 4:25  
136.1 C 222kPa 4:55  
136.2 C 222kPa 5:25  
136.4 C 221kPa 5:55  
136.4 C 223kPa 6:25  
136.2 C 221kPa 6:55  
136.1 C 222kPa 7:25  
136.1 C 222kPa 7:25  
VENTING 7:25  
AIR DRYING 8:18  
CYCLE COMPLETE 8:29

-----  
Digital Signature #  
52361271EE2D5B95  
-----