
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006061
9:17 02/08/2023

WRAPPED

135 C for 10 min.

CYCLE START	0:00
CONDITIONING	2:42
PRESSURIZING	4:41
STERILIZING	5:37
136.0 C 220kPa	5:37
136.4 C 224kPa	6:07
136.4 C 223kPa	6:37
136.1 C 219kPa	7:07
136.4 C 225kPa	7:37
136.3 C 221kPa	8:07
136.2 C 223kPa	8:37
136.4 C 223kPa	9:07
135.9 C 221kPa	9:37
136.3 C 221kPa	10:07
136.3 C 223kPa	10:37
136.3 C 224kPa	11:07
136.4 C 223kPa	11:37
136.4 C 223kPa	12:07
136.3 C 224kPa	12:37
136.4 C 224kPa	13:07
136.4 C 223kPa	13:37
136.4 C 223kPa	14:07
136.2 C 222kPa	14:37
136.3 C 224kPa	15:07
136.2 C 220kPa	15:37
136.2 C 220kPa	15:37
VENTING	15:37
AIR DRYING	16:27
CYCLE COMPLETE	20:14

Digital Signature #

EA207B4B5DEE6256
