
Midmark M11 - v1.0.0
Total Cycles: 3642

Name: RITTER 1
Sterilizer ID: V2178198
08/14/2023 10:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.4 F 0.0
2:00 165.6 F 2.4
4:00 166.7 F 5.6
6:00 196.2 F 9.7
8:00 216.1 F 6.6
10:00 224.3 F 7.8
12:00 232.2 F 12.6
14:00 235.2 F 9.3
16:00 240.7 F 10.0
18:00 246.4 F 13.5
20:00 257.7 F 19.6
22:00 268.0 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 271.8 F 28.5
2:00 272.2 F 28.8
3:00 272.0 F 29.1
4:00 271.1 F 28.3

Min 271.1 F 28.3
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 219.7 F 0.0
5:00 228.4 F 0.0
10:00 224.1 F 0.0
15:00 223.2 F 0.0
20:00 228.8 F 0.0
25:00 224.3 F 0.0
30:00 217.3 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 23:16
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:01:57
CYCLE COMPLETE

08/14/2023 11:21 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____