
Midmark M11 - v1.0.0
Total Cycles: 679

Name: MIDMARK
Sterilizer ID: V2178198
02/19/2021 07:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.0 F 0.0
2:00 78.7 F 0.0
4:00 106.2 F 0.0
6:00 114.7 F 0.1
8:00 139.9 F 0.1
10:00 175.5 F 0.2
12:00 192.7 F 0.6
14:00 205.4 F 3.2
16:00 214.6 F 5.5
18:00 222.8 F 8.2
20:00 231.4 F 11.5
22:00 228.8 F 6.6
24:00 240.3 F 11.3
26:00 236.9 F 8.9
28:00 242.5 F 10.8
30:00 247.3 F 13.5
32:00 258.1 F 19.3
34:00 268.0 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.9
1:00 272.5 F 28.6
2:00 272.3 F 28.3
3:00 271.6 F 27.8
4:00 272.1 F 28.4

Min 271.5 F 27.7
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.0 F 0.0
5:00 207.2 F 0.0
10:00 220.1 F 0.0
15:00 209.9 F 0.0
20:00 213.9 F 0.0
25:00 215.5 F 0.0
30:00 226.1 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 35:27
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:14:12
CYCLE COMPLETE

02/19/2021 09:02 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____