
Midmark M11 - v1.0.5
Total Cycles: 685

Name: RITTER2
Sterilizer ID: V2313228
01/12/2022 10:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 116.3 F 0.0
2:00 119.0 F 1.2
4:00 131.0 F 3.4
6:00 158.8 F 6.8
8:00 184.2 F 11.0
10:00 209.4 F 3.7
12:00 217.6 F 8.8
14:00 230.8 F 11.5
16:00 235.7 F 10.9
18:00 234.9 F 9.0
20:00 248.8 F 15.4
22:00 260.1 F 21.7
24:00 270.6 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.4
1:00 272.5 F 30.0
2:00 271.3 F 29.2
3:00 272.3 F 30.1
4:00 271.9 F 29.5

Min 271.2 F 29.2
Max 272.5 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.8 F 0.0
5:00 210.8 F 0.0
10:00 218.9 F 0.0
15:00 215.8 F 0.0
20:00 209.9 F 0.0
25:00 215.9 F 0.0
30:00 221.1 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 24:53
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:02:52
CYCLE COMPLETE

01/12/2022 11:05 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____