
Midmark M11 - v1.0.0
Total Cycles: 1991

Name: RITTER 1
Sterilizer ID: V2178198
06/06/2022 07:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.3 F 0.0
2:00 80.1 F 0.0
4:00 94.4 F 1.0
6:00 123.0 F 3.4
8:00 149.9 F 6.3
10:00 172.6 F 9.7
12:00 189.1 F 13.4
14:00 206.5 F 4.4
16:00 217.7 F 7.9
18:00 227.8 F 11.6
20:00 227.9 F 6.6
22:00 239.3 F 11.7
24:00 236.3 F 8.9
26:00 249.2 F 15.0
28:00 260.0 F 21.0
30:00 270.0 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.2 F 28.8
2:00 272.1 F 29.2
3:00 271.5 F 28.3
4:00 272.3 F 29.0

Min 271.3 F 28.3
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.5 F 0.0
5:00 204.7 F 0.0
10:00 223.6 F 0.0
15:00 226.1 F 0.0
20:00 222.2 F 0.0
25:00 222.6 F 0.0
30:00 225.8 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 30:55
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:08:48
CYCLE COMPLETE

06/06/2022 08:33 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____