
Midmark M11 - v1.0.5
Total Cycles: 4121

Name: RITTER2
Sterilizer ID: V2313228
11/14/2024 04:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.4 F 0.0
2:00 140.7 F 0.1
4:00 149.5 F 2.4
6:00 178.2 F 6.0
8:00 196.6 F 10.4
10:00 220.0 F 6.3
12:00 228.1 F 10.1
14:00 233.2 F 8.5
16:00 239.5 F 12.8
18:00 240.8 F 11.6
20:00 253.4 F 17.9
22:00 264.0 F 24.5
24:00 271.7 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.9
1:00 272.4 F 29.9
2:00 272.7 F 30.3
3:00 272.3 F 30.5
4:00 271.9 F 29.8

Min 271.7 F 29.8
Max 272.8 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.0 F 0.0
5:00 232.1 F 0.0
10:00 228.6 F 0.0
15:00 226.2 F 0.0
20:00 221.4 F 0.0
Dry Cycle Aborted

FILLING: 1:03
HEATING: 24:02
STERILIZING: 4:00
VENTING: 3:33
DRYING: 24:42
TOTAL CYCLE: 00:57:20

11/14/2024 05:47 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____