
Midmark M11 - v1.0.0
Total Cycles: 622

Name: MIDMARK
Sterilizer ID: V2178198
01/25/2021 07:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.5 F 0.0
2:00 71.7 F 0.0
4:00 75.3 F 0.0
6:00 94.5 F 0.1
8:00 125.0 F 0.2
10:00 171.3 F 0.3
12:00 201.6 F 3.5
14:00 213.8 F 6.7
16:00 224.0 F 10.1
18:00 229.6 F 9.8
20:00 238.9 F 9.6
22:00 238.1 F 9.6
24:00 246.2 F 12.9
26:00 243.6 F 11.7
28:00 257.0 F 18.6
30:00 268.3 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 27.8
1:00 272.4 F 28.2
2:00 272.2 F 28.4
3:00 271.8 F 27.8
4:00 272.7 F 28.6

Min 271.4 F 27.6
Max 272.7 F 28.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.8 F 0.0
5:00 248.7 F 0.0
10:00 227.5 F 0.0
15:00 227.2 F 0.0
20:00 224.0 F 0.0
25:00 226.7 F 0.0
30:00 223.7 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 31:11
STERILIZING: 4:00
VENTING: 2:58
DRYING: 30:00
TOTAL CYCLE: 01:09:00
CYCLE COMPLETE

01/25/2021 08:45 AM
ACCEPTED: X
REJECTED:
Operator: U01