
Midmark M11 - v1.0.5
Total Cycles: 3765

Name: RITTER2
Sterilizer ID: V2313228
07/10/2024 02:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 135.2 F 0.0
2:00 135.4 F 1.6
4:00 146.2 F 4.5
6:00 172.8 F 8.1
8:00 189.9 F 12.2
10:00 214.4 F 5.5
12:00 222.8 F 9.6
14:00 231.5 F 10.6
16:00 237.1 F 10.9
18:00 234.0 F 8.7
20:00 248.0 F 14.9
22:00 258.7 F 21.0
24:00 268.9 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.8
1:00 272.0 F 29.8
2:00 271.9 F 29.7
3:00 272.2 F 29.7
4:00 272.4 F 29.9

Min 271.8 F 29.7
Max 273.0 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.4 F 0.0
5:00 216.7 F 0.0
10:00 214.9 F 0.0
15:00 196.1 F 0.0
20:00 190.8 F 0.0
25:00 187.5 F 0.0
30:00 187.0 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 25:50
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:04:37
CYCLE COMPLETE

07/10/2024 03:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____