
Midmark M11 - v1.0.0
Total Cycles: 3805

Name: RITTER 1
Sterilizer ID: V2178198
09/21/2023 11:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.7 F 0.0
2:00 175.7 F 2.1
4:00 191.3 F 6.4
6:00 211.5 F 11.6
8:00 219.3 F 6.3
10:00 232.2 F 11.4
12:00 233.7 F 8.4
14:00 241.0 F 11.0
16:00 245.5 F 12.9
18:00 257.4 F 19.2
20:00 268.1 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.8 F 28.9
2:00 271.6 F 28.2
3:00 272.8 F 29.0
4:00 271.9 F 28.3

Min 271.4 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.3 F 0.0
5:00 204.9 F 0.0
10:00 211.4 F 0.0
15:00 219.4 F 0.0
20:00 223.9 F 0.0
25:00 229.2 F 0.0
30:00 231.9 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 21:20
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:00:17
CYCLE COMPLETE

09/21/2023 12:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____