
Midmark M11 - v1.0.0
Total Cycles: 2703

Name: RITTER 1
Sterilizer ID: V2178198
12/20/2022 04:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.8 F 0.0
2:00 206.5 F 4.3
4:00 216.2 F 9.9
6:00 225.5 F 5.7
8:00 235.3 F 11.2
10:00 236.0 F 8.8
12:00 240.7 F 9.5
14:00 248.9 F 14.7
16:00 261.7 F 21.9
18:00 272.7 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.4 F 28.8
2:00 272.3 F 29.1
3:00 272.0 F 28.6
4:00 272.4 F 29.2

Min 271.5 F 28.3
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 216.4 F 0.0
5:00 227.6 F 0.0
10:00 236.0 F 0.0
15:00 234.7 F 0.0
20:00 226.5 F 0.0
25:00 225.5 F 0.0
30:00 230.6 F 0.0
Drying Complete

FILLING: 2:01
HEATING: 18:17
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:31
CYCLE COMPLETE

12/20/2022 05:40 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____