
Midmark M11 - v1.0.5
Total Cycles: 4224

Name: RITTER2
Sterilizer ID: V2313228
12/18/2024 10:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.2 F 0.0
2:00 73.7 F 0.0
4:00 107.1 F 0.0
6:00 137.1 F 0.1
8:00 148.3 F 0.7
10:00 176.9 F 3.7
12:00 194.5 F 6.7
14:00 209.1 F 10.5
16:00 222.5 F 7.0
18:00 227.5 F 8.0
20:00 237.6 F 12.3
22:00 235.3 F 9.2
24:00 240.1 F 10.4
26:00 246.0 F 13.4
28:00 257.4 F 19.6
30:00 267.9 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.8 F 29.4
2:00 271.7 F 28.6
3:00 272.9 F 29.5
4:00 272.1 F 28.8

Min 271.6 F 28.5
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.6 F 0.0
5:00 199.4 F 0.0
10:00 194.9 F 0.0
15:00 189.8 F 0.0
20:00 202.5 F 0.0
25:00 213.5 F 0.0
30:00 226.3 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 31:24
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:09:36
CYCLE COMPLETE

12/18/2024 11:54 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____