
Midmark M11 - v1.0.5
Total Cycles: 4149

Name: RITTER2
Sterilizer ID: V2313228
11/21/2024 02:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 129.6 F 0.0
2:00 129.3 F 0.0
4:00 143.6 F 0.9
6:00 173.7 F 4.1
8:00 199.5 F 7.6
10:00 209.6 F 11.9
12:00 221.0 F 6.5
14:00 231.0 F 11.2
16:00 231.7 F 8.4
18:00 241.0 F 11.8
20:00 243.8 F 13.1
22:00 256.0 F 19.1
24:00 266.3 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.1
1:00 272.4 F 29.9
2:00 272.9 F 30.2
3:00 272.0 F 30.1
4:00 272.6 F 30.2

Min 270.8 F 28.9
Max 273.3 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.2 F 0.0
5:00 219.1 F 0.0
10:00 219.9 F 0.0
15:00 209.8 F 0.0
20:00 204.2 F 0.0
25:00 201.4 F 0.0
30:00 200.4 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 25:42
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:04:25
CYCLE COMPLETE

11/21/2024 03:31 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____