
Midmark M11 - v1.0.5
Total Cycles: 2874

Name: RITTER2
Sterilizer ID: V2313228
09/12/2023 05:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.9 F 0.0
2:00 185.7 F 4.5
4:00 207.2 F 10.3
6:00 221.2 F 6.3
8:00 237.2 F 12.5
10:00 239.1 F 10.9
12:00 239.0 F 10.2
14:00 254.3 F 17.9
16:00 267.1 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.1 F 28.9
2:00 272.7 F 29.5
3:00 271.6 F 28.6
4:00 272.5 F 29.2

Min 271.5 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.4 F 0.0
5:00 222.0 F 0.0
10:00 229.0 F 0.0
15:00 228.6 F 0.0
20:00 222.6 F 0.0
25:00 221.1 F 0.0
30:00 218.7 F 0.0
Drying Complete

FILLING: 2:46
HEATING: 17:17
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:57:12
CYCLE COMPLETE

09/12/2023 06:00 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____