
Midmark M11 - v1.0.5
Total Cycles: 3261

Name: RITTER2
Sterilizer ID: V2313228
12/28/2023 12:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.3 F 0.0
2:00 146.0 F 1.9
4:00 170.1 F 5.5
6:00 194.1 F 10.3
8:00 212.4 F 5.1
10:00 228.7 F 9.4
12:00 236.2 F 9.5
14:00 241.3 F 12.1
16:00 239.8 F 10.7
18:00 253.2 F 17.4
20:00 264.8 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.1 F 29.1
2:00 272.2 F 29.6
3:00 271.0 F 28.8
4:00 272.5 F 30.0

Min 271.0 F 28.8
Max 273.1 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.2 F 0.0
5:00 228.9 F 0.0
10:00 208.0 F 0.0
15:00 185.1 F 0.0
20:00 181.6 F 0.0
25:00 180.2 F 0.0
30:00 178.9 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 21:48
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:00:01
CYCLE COMPLETE

12/28/2023 01:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____