

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
2.3uS / 1.4ppm  
CYCLE NUMBER 001867  
8:02 18/07/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:13  
PRESSURIZING 2:18  
STERILIZING 3:26  
136.4 C 224kPa 3:26  
136.3 C 225kPa 3:56  
136.4 C 225kPa 4:26  
136.2 C 224kPa 4:56  
136.4 C 222kPa 5:26  
136.2 C 223kPa 5:56  
136.4 C 224kPa 6:26  
136.4 C 222kPa 6:56  
136.4 C 222kPa 6:56  
VENTING 6:56  
AIR DRYING 7:48  
CYCLE COMPLETE 23:18  
-----

Digital Signature #  
6BDF610C6129F0CB  
-----