
Midmark M11 - v1.0.0
Total Cycles: 4703

Name: RITTER 1
Sterilizer ID: V2178198
07/08/2024 08:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.1 F 0.0
2:00 68.7 F 0.3
4:00 88.9 F 1.3
6:00 119.6 F 3.3
8:00 140.0 F 6.0
10:00 166.4 F 9.3
12:00 184.5 F 13.0
14:00 207.3 F 3.5
16:00 215.7 F 6.7
18:00 226.0 F 10.2
20:00 232.3 F 9.3
22:00 236.7 F 9.9
24:00 238.2 F 9.3
26:00 244.9 F 12.7
28:00 256.6 F 18.8
30:00 267.0 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.5 F 28.9
2:00 271.4 F 28.3
3:00 272.5 F 28.9
4:00 271.4 F 28.1

Min 271.4 F 28.1
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.3 F 0.0
5:00 204.3 F 0.0
10:00 216.5 F 0.0
15:00 215.5 F 0.0
20:00 210.2 F 0.0
25:00 210.2 F 0.0
30:00 212.7 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 31:33
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:09:46
CYCLE COMPLETE

07/08/2024 09:59 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____