
Midmark M11 - v1.0.5
Total Cycles: 1791

Name: RITTER2
Sterilizer ID: V2313228
11/30/2022 10:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 204.7 F 0.0
2:00 196.1 F 2.1
4:00 205.9 F 7.9
6:00 223.7 F 11.5
8:00 232.6 F 10.6
10:00 234.1 F 9.0
12:00 237.6 F 9.6
14:00 250.5 F 16.2
16:00 263.4 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.2
1:00 272.3 F 29.6
2:00 271.9 F 29.8
3:00 272.1 F 29.4
4:00 272.0 F 29.9

Min 271.4 F 29.1
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.7 F 0.0
5:00 227.6 F 0.0
10:00 232.8 F 0.0
15:00 228.2 F 0.0
20:00 228.2 F 0.0
25:00 227.8 F 0.0
30:00 226.5 F 0.0

Drying Complete

FILLING: 1:33
HEATING: 17:57
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:56:41
CYCLE COMPLETE

11/30/2022 11:41 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____