
Midmark M11 - v1.0.5
Total Cycles: 2022

Name: RITTER2
Sterilizer ID: V2313228
02/08/2023 08:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 77.9 F 0.0
2:00 81.9 F 0.0
4:00 116.0 F 0.0
6:00 122.0 F 0.0
8:00 168.0 F 0.0
10:00 180.9 F 2.3
12:00 197.7 F 5.2
14:00 212.5 F 8.8
16:00 225.4 F 13.0
18:00 230.3 F 8.5
20:00 241.5 F 12.8
22:00 241.3 F 11.7
24:00 240.0 F 10.6
26:00 253.3 F 17.3
28:00 264.8 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.5 F 29.1
2:00 271.6 F 28.9
3:00 272.4 F 29.1
4:00 271.5 F 28.7

Min 271.4 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.8 F 0.0
5:00 207.1 F 0.0
10:00 212.9 F 0.0
15:00 205.7 F 0.0
20:00 208.8 F 0.0
25:00 219.9 F 0.0
30:00 226.9 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 29:51
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:07:47
CYCLE COMPLETE

02/08/2023 09:33 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____