
Midmark M11 - v1.0.5
Total Cycles: 3111

Name: RITTER2
Sterilizer ID: V2313228
11/10/2023 11:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.4 F 0.0
2:00 158.0 F 2.2
4:00 171.1 F 5.8
6:00 194.5 F 10.5
8:00 212.1 F 5.2
10:00 224.7 F 9.6
12:00 231.2 F 9.1
14:00 240.4 F 12.4
16:00 240.5 F 11.3
18:00 253.7 F 18.0
20:00 265.4 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.6 F 30.0
2:00 272.2 F 29.5
3:00 271.3 F 29.2
4:00 272.3 F 30.1

Min 271.2 F 29.1
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.2 F 0.0
5:00 210.2 F 0.0
10:00 218.5 F 0.0
15:00 207.8 F 0.0
20:00 206.7 F 0.0
25:00 213.8 F 0.0
30:00 227.8 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 21:50
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:00:21
CYCLE COMPLETE

11/10/2023 12:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____