
Midmark M11 - v1.0.0
Total Cycles: 4702

Name: RITTER 1
Sterilizer ID: V2178198
07/05/2024 03:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 94.5 F 0.0
2:00 98.2 F 0.9
4:00 121.5 F 3.1
6:00 151.8 F 5.8
8:00 174.8 F 9.1
10:00 200.3 F 13.0
12:00 209.0 F 5.0
14:00 220.9 F 8.6
16:00 230.7 F 12.8
18:00 231.9 F 8.5
20:00 243.2 F 13.5
22:00 241.4 F 11.4
24:00 253.1 F 17.2
26:00 263.6 F 23.3
28:00 272.3 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.5
1:00 272.4 F 29.3
2:00 271.7 F 28.6
3:00 271.3 F 28.8
4:00 272.6 F 29.3

Min 271.0 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.2 F 0.0
5:00 203.0 F 0.0
10:00 194.2 F 0.0
15:00 189.5 F 0.0
20:00 191.4 F 0.0
25:00 195.9 F 0.0
30:00 200.1 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 28:16
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:06:53
CYCLE COMPLETE

07/05/2024 04:08 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____