
Midmark M11 - v1.0.0
Total Cycles: 3140

Name: RITTER 1
Sterilizer ID: V2178198
04/20/2023 03:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.2 F 0.0
2:00 194.3 F 0.0
4:00 214.3 F 5.1
6:00 229.0 F 10.4
8:00 230.9 F 7.2
10:00 243.5 F 13.1
12:00 242.8 F 11.7
14:00 241.8 F 10.9
16:00 256.0 F 18.2
18:00 268.1 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 272.5 F 28.5
2:00 271.6 F 28.1
3:00 272.6 F 28.6
4:00 271.6 F 27.9

Min 271.5 F 27.9
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.1 F 0.0
5:00 218.8 F 0.0
10:00 238.4 F 0.0
15:00 233.3 F 0.0
20:00 226.0 F 0.0
25:00 223.6 F 0.0
30:00 224.0 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 19:13
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:58:10
CYCLE COMPLETE

04/20/2023 04:03 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____