

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 003822  
12:08 15/06/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:56  
PRESSURIZING 2:00  
STERILIZING 2:57  
136.4 C 220kPa 2:57  
136.4 C 223kPa 3:27  
136.4 C 223kPa 3:57  
136.2 C 224kPa 4:27  
136.4 C 224kPa 4:57  
136.4 C 225kPa 5:27  
136.2 C 221kPa 5:57  
136.2 C 223kPa 6:27  
136.2 C 223kPa 6:27  
VENTING 6:27  
AIR DRYING 8:30  
CYCLE COMPLETE 9:40

-----  
Digital Signature #  
64CF48E91AAFAF2E  
-----