
Midmark M11 - v1.0.0
Total Cycles: 3265

Name: RITTER 1
Sterilizer ID: V2178198
05/23/2023 08:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.1 F 0.0
2:00 175.0 F 0.0
4:00 185.4 F 2.5
6:00 205.4 F 7.4
8:00 224.0 F 12.3
10:00 228.9 F 8.0
12:00 242.1 F 13.5
14:00 241.2 F 11.4
16:00 239.2 F 9.9
18:00 253.2 F 16.8
20:00 265.0 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.7 F 29.1
2:00 272.3 F 28.6
3:00 271.5 F 28.1
4:00 272.1 F 28.9

Min 271.4 F 28.1
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.3 F 0.0
5:00 229.7 F 0.0
10:00 240.5 F 0.0
15:00 233.6 F 0.0
20:00 225.8 F 0.0
25:00 225.7 F 0.0
30:00 225.4 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 21:56
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:00:32
CYCLE COMPLETE

05/23/2023 09:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____