
Midmark M11 - v1.0.5
Total Cycles: 1914

Name: RITTER2
Sterilizer ID: V2313228
01/10/2023 03:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.0 F 0.0
2:00 175.2 F 3.6
4:00 188.9 F 8.2
6:00 211.4 F 10.7
8:00 224.8 F 8.9
10:00 234.1 F 9.7
12:00 242.2 F 12.9
14:00 242.9 F 12.3
16:00 256.5 F 19.4
18:00 268.4 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.5 F 29.7
2:00 271.5 F 28.9
3:00 272.3 F 29.9
4:00 272.2 F 29.4

Min 271.2 F 28.9
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.2 F 0.0
5:00 232.3 F 0.0
10:00 243.0 F 0.0
15:00 236.3 F 0.0
20:00 227.5 F 0.0
25:00 227.5 F 0.0
30:00 231.1 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 19:12
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:07
CYCLE COMPLETE

01/10/2023 03:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____