
Midmark M11 - v1.0.5
Total Cycles: 1443

Name: RITTER2
Sterilizer ID: V2313228
08/31/2022 04:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.5 F 0.0
2:00 174.5 F 2.7
4:00 185.5 F 6.7
6:00 210.9 F 11.8
8:00 222.6 F 7.4
10:00 234.9 F 12.9
12:00 237.1 F 10.8
14:00 236.3 F 9.5
16:00 250.3 F 16.5
18:00 262.6 F 23.4
20:00 272.3 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.5
1:00 272.5 F 29.9
2:00 271.4 F 29.5
3:00 272.5 F 30.1
4:00 271.8 F 29.4

Min 271.4 F 29.3
Max 272.7 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.6 F 0.0
5:00 213.6 F 0.0
10:00 235.6 F 0.0
15:00 238.8 F 0.0
20:00 227.1 F 0.0
25:00 222.0 F 0.0
30:00 223.3 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 20:13
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:41
CYCLE COMPLETE

08/31/2022 05:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____