
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 004357
10:14 14/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:31
PRESSURIZING 2:34
STERILIZING 3:26
136.0 C 220kPa 3:26
136.3 C 224kPa 3:56
136.4 C 223kPa 4:26
136.4 C 221kPa 4:56
136.1 C 223kPa 5:26
136.3 C 224kPa 5:56
136.4 C 223kPa 6:26
136.0 C 221kPa 6:56
136.0 C 221kPa 6:56
VENTING 6:56
AIR DRYING 7:50
CYCLE COMPLETE 9:35

Digital Signature #
72F6E2EA2240C21E
