
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002112
7:38 03/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:45
PRESSURIZING 2:52
STERILIZING 4:04
136.0 C 220kPa 4:04
136.1 C 222kPa 4:34
136.2 C 221kPa 5:04
136.4 C 224kPa 5:34
136.3 C 221kPa 6:04
136.3 C 223kPa 6:34
136.2 C 222kPa 7:04
136.4 C 221kPa 7:34
136.4 C 221kPa 7:34
VENTING 7:34
AIR DRYING 8:27
CYCLE COMPLETE 38:27

Digital Signature #
3E7083D410C85E90
