
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 005492
9:43 04/05/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:07
PRESSURIZING 2:11
STERILIZING 3:11
136.4 C 221kPa 3:11
136.4 C 225kPa 3:41
136.2 C 223kPa 4:11
136.2 C 223kPa 4:41
136.4 C 224kPa 5:11
136.4 C 224kPa 5:41
136.0 C 220kPa 6:11
136.4 C 223kPa 6:41
136.4 C 223kPa 6:41
VENTING 6:41
AIR DRYING 7:30
CYCLE COMPLETE 12:41

Digital Signature #
EA35873E94CF3343
