
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000988
7:56 14/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:28
PRESSURIZING 2:36
STERILIZING 3:42
136.4 C 230kPa 3:42
136.4 C 232kPa 4:12
136.2 C 230kPa 4:42
136.4 C 231kPa 5:12
136.3 C 230kPa 5:42
136.2 C 231kPa 6:12
136.3 C 231kPa 6:42
136.4 C 231kPa 7:12
136.4 C 231kPa 7:12
VENTING 7:12
AIR DRYING 8:05
CYCLE COMPLETE 8:18

Digital Signature #
FF7BDD0F8D0F83C0
