
Midmark M11 - v1.0.0
Total Cycles: 2021

Name: RITTER 1
Sterilizer ID: V2178198
06/13/2022 08:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.0 F 0.0
2:00 178.7 F 2.3
4:00 184.3 F 6.5
6:00 206.8 F 12.0
8:00 221.9 F 6.5
10:00 236.9 F 12.4
12:00 238.0 F 9.8
14:00 236.2 F 8.1
16:00 249.8 F 14.8
18:00 262.1 F 22.0
20:00 272.4 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.0 F 28.1
2:00 272.6 F 28.6
3:00 272.8 F 28.9
4:00 271.9 F 28.4

Min 271.5 F 27.9
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 224.4 F 0.0
5:00 223.4 F 0.0
10:00 233.1 F 0.0
15:00 225.4 F 0.0
20:00 219.6 F 0.0
Dry Cycle Aborted

FILLING: 1:59
HEATING: 20:13
STERILIZING: 4:00
VENTING: 2:48
DRYING: 23:05
TOTAL CYCLE: 00:52:05

06/13/2022 09:30 AM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____