

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5181

Name: RITTER 1  
Sterilizer ID: V2178198  
12/12/2024 10:00 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.5 F 0.0  
2:00 72.8 F 0.0  
4:00 85.5 F 0.0  
6:00 122.8 F 0.0  
8:00 171.8 F 0.1  
10:00 187.2 F 0.2  
12:00 206.8 F 1.0  
14:00 224.7 F 4.4  
16:00 231.3 F 7.1  
18:00 239.4 F 10.4  
20:00 234.2 F 8.9  
22:00 242.4 F 11.4  
24:00 237.0 F 9.0  
26:00 238.9 F 9.4  
28:00 248.2 F 14.1  
30:00 260.3 F 20.8  
32:00 271.1 F 27.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.0  
1:00 272.8 F 28.8  
2:00 271.8 F 27.9  
3:00 272.5 F 28.7  
4:00 272.4 F 28.4  
  
Min 271.6 F 27.9  
Max 272.9 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.4 F 0.0  
5:00 232.7 F 0.0  
10:00 204.4 F 0.0  
15:00 204.8 F 0.0  
20:00 198.0 F 0.0  
25:00 209.3 F 0.0  
30:00 201.9 F 0.0  
  
Drying Complete

FILLING: 0:40  
HEATING: 32:45  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 01:10:44  
CYCLE COMPLETE

12/12/2024 11:10 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_