
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 008102
10:47 27/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:44
PRESSURIZING 2:48
STERILIZING 3:49
136.4 C 220kPa 3:49
136.1 C 223kPa 4:19
136.4 C 223kPa 4:49
136.1 C 223kPa 5:19
136.4 C 221kPa 5:49
135.9 C 220kPa 6:19
136.4 C 224kPa 6:49
136.4 C 224kPa 7:19
136.4 C 224kPa 7:19
VENTING 7:19
AIR DRYING 8:08
CYCLE COMPLETE 8:31

Digital Signature #
8A5CF4CC5290AB32
