
Midmark M11 - v1.0.0
Total Cycles: 5405

Name: RITTER 1
Sterilizer ID: V2178198
03/10/2025 05:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 118.9 F 0.0
2:00 139.7 F 1.3
4:00 155.2 F 4.3
6:00 178.5 F 7.8
8:00 192.3 F 11.5
10:00 213.2 F 2.8
12:00 217.1 F 6.7
14:00 229.0 F 10.6
16:00 231.7 F 7.4
18:00 240.0 F 11.5
20:00 236.8 F 9.0
22:00 249.6 F 15.0
24:00 260.4 F 21.1
26:00 270.5 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.3
1:00 272.4 F 29.4
2:00 272.3 F 29.4
3:00 272.1 F 29.4
4:00 271.6 F 29.1

Min 270.6 F 28.2
Max 273.1 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.6 F 0.0
5:00 204.9 F 0.0
10:00 185.8 F 0.0
15:00 231.4 F 0.0
20:00 227.9 F 0.0
25:00 234.0 F 0.0
30:00 229.2 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 26:45
STERILIZING: 4:00
VENTING: 3:52
DRYING: 30:00
TOTAL CYCLE: 01:05:28
CYCLE COMPLETE

03/10/2025 06:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____