

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3512

Name: RITTER 1  
Sterilizer ID: V2178198  
07/11/2023 11:25 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 186.4 F 0.0  
2:00 185.7 F 1.1  
4:00 198.2 F 5.1  
6:00 218.6 F 10.3  
8:00 224.3 F 6.1  
10:00 236.4 F 11.1  
12:00 234.9 F 8.5  
14:00 240.8 F 10.3  
16:00 247.1 F 13.8  
18:00 259.4 F 20.5  
20:00 270.5 F 27.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.3  
1:00 272.4 F 28.7  
2:00 271.7 F 28.7  
3:00 272.1 F 28.6  
4:00 271.7 F 28.8  
  
Min 271.3 F 28.2  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.5 F 0.0  
5:00 212.9 F 0.0  
10:00 226.1 F 0.0  
15:00 230.1 F 0.0  
20:00 228.0 F 0.0  
25:00 227.2 F 0.0  
30:00 224.4 F 0.0  
Drying Complete

FILLING: 1:31  
HEATING: 20:48  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:59:35  
CYCLE COMPLETE

07/11/2023 12:24 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_