
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000375
12:54 21/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:57
PRESSURIZING 3:04
STERILIZING 4:16
136.0 C 223kPa 4:16
136.3 C 233kPa 4:46
136.4 C 231kPa 5:16
136.2 C 232kPa 5:46
136.4 C 233kPa 6:16
136.3 C 232kPa 6:46
136.4 C 233kPa 7:16
136.4 C 231kPa 7:46
136.4 C 231kPa 7:46
VENTING 7:46
AIR DRYING 8:40
CYCLE COMPLETE 8:46

Digital Signature #
48C015EABDA16C68
