
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 007854
14:11 10/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:02
STERILIZING 4:16
136.0 C 220kPa 4:16
136.3 C 223kPa 4:46
136.3 C 222kPa 5:16
136.0 C 220kPa 5:46
136.1 C 221kPa 6:16
136.5 C 224kPa 6:46
136.1 C 222kPa 7:16
136.3 C 224kPa 7:46
136.3 C 224kPa 7:46
VENTING 7:46
AIR DRYING 8:39
CYCLE COMPLETE 38:39

Digital Signature #
8BF190B5C22A6FF2
