
Midmark M11 - v1.0.0
Total Cycles: 2376

Name: RITTER 1
Sterilizer ID: V2178198
09/27/2022 09:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.9 F 0.0
2:00 194.1 F 3.7
4:00 208.9 F 9.1
6:00 226.8 F 6.5
8:00 232.0 F 10.8
10:00 235.3 F 7.9
12:00 241.6 F 10.9
14:00 245.7 F 13.3
16:00 258.8 F 20.4
18:00 270.4 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.3 F 29.3
2:00 271.7 F 28.6
3:00 271.8 F 29.1
4:00 272.3 F 29.2

Min 271.2 F 28.4
Max 272.3 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.1 F 0.0
5:00 214.6 F 0.0
10:00 227.6 F 0.0
15:00 230.8 F 0.0
20:00 228.5 F 0.0
25:00 229.2 F 0.0
30:00 231.2 F 0.0
Drying Complete

FILLING: 2:30
HEATING: 18:50
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:35
CYCLE COMPLETE

09/27/2022 10:45 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____