
Midmark M11 - v1.0.0
Total Cycles: 3790

Name: RITTER 1
Sterilizer ID: V2178198
09/18/2023 01:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.5 F 0.0
2:00 169.7 F 2.2
4:00 190.7 F 6.2
6:00 208.4 F 10.6
8:00 226.9 F 5.9
10:00 232.4 F 10.2
12:00 235.1 F 8.3
14:00 244.7 F 13.1
16:00 243.5 F 11.9
18:00 255.7 F 18.3
20:00 266.6 F 24.9
22:00 272.9 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 271.9 F 28.6
2:00 272.2 F 28.9
3:00 272.7 F 29.2
4:00 271.6 F 28.7

Min 271.5 F 28.5
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.4 F 0.0
5:00 214.6 F 0.0
10:00 191.8 F 0.0
15:00 189.9 F 0.0
20:00 187.9 F 0.0
25:00 185.1 F 0.0
30:00 190.7 F 0.0
Drying Complete

FILLING: 1:28
HEATING: 22:14
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:01:23
CYCLE COMPLETE

09/18/2023 02:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____